

Product Specification Sheet

Delvotest® Incubator MiniS

20240527

Description

The Delvotest® dry incubator with integrated thermometer and timer functions has been developed for incubating the antibiotic residue test ampoule presentations of Delvotest®. It is a ready-to-use electronic heating device with a thermostatically controlled temperature, which meets the required specifications for optimal performance of Delvotest® ampoules. The incubator is presented in a plastic casing with an internal metal incubating block for positioning 10 Delvotest® ampoules. The dry incubator has a digital thermometer and timer functions programmable to customer needs. The device comes with a Continental European CEE VIIG plug (some countries may have other plugs appropriate for local use). A cigarette lighter plug adapter 12/24 Vdc is available upon request. Verify country regulations on use of accessories while driving.

Technical data

Thermometer and timer settings:

- Temperature range: RT+ 5 ~ 80 °C
- Temperature accuracy: $\leq \pm 0.5$ °C (Contact your local DSM representative for more information on the temperature accuracy validation method)
- Display accuracy: ± 0.1 °C
- Heating time (from 25°C to 70°C): ≤ 20 min
- Timing range: 1h ~ 999h
1min ~ 999min
1sec ~ 999sec
- Operating temperature (Surrounding): + 5 to + 35 °C

Timer settings:

- Alarm sound or electronic switch-off setting
- Time set programmable (hh:mm): PrA (00:07 at 64°C), PrB (03:15 at 64°C), PrC (00:02 at 64°C), PrD (00:03 at 64°C), PrE (no preset time/temp)

Technical data:

Power supply: DC12V 1.5 A
Power Capacity: 15 W
Average power Consumption: 5 W
Weight: 0.5 kg
Dimensions: 99 x 135 x 34 mm (W x D x H)
Certification: Low Voltage Directive 2014/35/EU, Electromagnetic Compatibility 2014/30/EU, EN 61010-1:2010+A1:2019 CE-conformity: EN 61326-1:2013

Warranty:

Unless stipulated otherwise in the relevant sales agreement, DSM provides a guarantee on the Incubator, which guarantee is stipulated in the General terms and Conditions of Sale of DSM ("GTCS"). Supported Accessories are explicitly excluded from any guarantee given by DSM. The Incubator is guaranteed to be free from defective materials and workmanship within the defined period, provided that misuse, alteration and/or abnormal use have not occurred by user. It will be repaired under warranty for a period of 2 (two) years from the date of manufacture printed on the bottom of the incubator casing. Do not open or repair defective incubators but contact DSM representative immediately when a defect occurs. Products NOT repaired by DSM (third party repair) are explicitly excluded from any guarantee or warranty given by DSM. In case of contradictions between these specifications and the GTCS, the GTCS shall prevail.

Although diligent care has been used to ensure that the information provided herein is accurate, nothing contained herein can be construed to imply any representation or warranty for which we assume legal responsibility, including without limitation any warranties as to the accuracy, currency or completeness of this information or of noninfringement of third party intellectual property rights. The content of this document is subject to change without further notice. This document is non-controlled and will not be automatically replaced when changed. Please contact us for the latest version of this document or for further information. Since the user's product formulations, specific use applications and conditions of use are beyond our control, we make no warranty or representation regarding the results which may be obtained by the user. It shall be the responsibility of the user to determine the suitability of our products for the user's specific purposes and the legal status for the user's intended use of our products.

©DSM Food Specialties B.V. | Trade Register Number 27235314