

CERTIFICATE OF QUALITY ASSURANCE

Description : Measuring Cylinder
Material : Polypropylene
Catalog Number : 016.05.250
Visual appearance : Within acceptability limits

Volume	Diameter	Height	Thickness
ml	mm	mm	mm
250	40 (± 1.0)	320 (± 3.0)	4,4 – 5,1

The above mentioned properties tested by ISOLAB Quality Control Laboratory.

For and on behalf of

Quality Assurance Department
ISOLAB LABORGERÄTE GmbH

Date

07.04.2016

Quality Assurance Manager

